Food and Drug Administration, HHS

§882.1835 Physiological signal amplifier.

- (a) *Identification*. A physiological signal amplifier is a general purpose device used to electrically amplify signals derived from various physiological sources (e.g., the electroencephalogram).
- (b) Classification. Class II (performance standards).

§882.1845 Physiological signal conditioner.

- (a) *Identification*. A physiological signal conditioner is a device such as an integrator or differentiator used to modify physiological signals for recording and processing.
- (b) Classification. Class II (performance standards).

§ 882.1855 Electroencephalogram (EEG) telemetry system.

- (a) *Identification*. An electroencephalogram (EEG) telemetry system consists of transmitters, receivers, and other components used for remotely monitoring or measuring EEG signals by means of radio or telephone transmission systems.
- (b) Classification. Class II (performance standards).

§ 882.1870 Evoked response electrical stimulator.

- (a) *Identification*. An evoked response electrical stimulator is a device used to apply an electrical stimulus to a patient by means of skin electrodes for the purpose of measuring the evoked response.
- (b) Classification. Class II (performance standards).

§882.1880 Evoked response mechanical stimulator.

- (a) *Identification*. An evoked response mechanical stimulator is a device used to produce a mechanical stimulus or a series of mechanical stimuli for the purpose of measuring a patient's evoked response.
- (b) Classification. Class II (performance standards).

§ 882.1890 Evoked response photic stimulator.

(a) *Identification*. An evoked response photic stimulator is a device used to

generate and display a shifting pattern or to apply a brief light stimulus to a patient's eye for use in evoked response measurements or for electroencephalogram (EEG) activation.

(b) Classification. Class II (performance standards).

§882.1900 Evoked response auditory stimulator.

- (a) *Identification*. An evoked response auditory stimulator is a device that produces a sound stimulus for use in evoked response measurements or electroencephalogram activation.
- (b) Classification. Class II (performance standards).

§882.1925 Ultrasonic scanner calibration test block.

- (a) *Identification*. An ultrasonic scanner calibration test block is a block of material with known properties used to calibrate ultrasonic scanning devices (e.g., the echoencephalograph).
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §882.9.

[44 FR 51730-51778, Sept. 4, 1979, as amended at 59 FR 63011, Dec. 7, 1994; 66 FR 38807, July 25, 20011

§882.1950 Tremor transducer.

- (a) *Identification*. A tremor transducer is a device used to measure the degree of tremor caused by certain diseases.
- (b) Classification. Class II (performance standards).

Subparts C-D [Reserved]

Subpart E—Neurological Surgical Devices

§882.4030 Skull plate anvil.

- (a) *Identification*. A skull plate anvil is a device used to form alterable skull plates in the proper shape to fit the curvature of a patient's skull.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in